

# Display Report

## Analysis Info

Analysis Name D:\data\fenglei\CS018158\_BD4\_01\_26651.d  
Method 1008-MID.m  
Sample Name CS018158  
Comment

Acquisition Date 2/27/2024 6:27:00 PM  
Operator Demo User  
Instrument impact II 1825265.10256

## Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.3 Bar
Focus	Active	Set Capillary	3500 V	Set Dry Heater	200 °C
Scan Begin	150 m/z	Set End Plate Offset	-500 V	Set Dry Gas	3.0 l/min
Scan End	8000 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

